

ADDITIONAL SPECIFICATIONS

FDA code	MMC Item code	Product type	Size	Sterile	Specs	Certification	DOH approval date
		Disinfectant wipes	Pks of 100		Must be listed on current EPA registration List N	EPA List N https://www.epa.gov/pesticide-registration/list-n-disinfectants-use-against-sars-cov-2	April 1 2020
		Face shield			Prefer ANSI Z87.1 but not required. https://www.fda.gov/media/136842/download	ANSI Z87.1	April 1 2020
OPJ		Gloves, exam	S, M, L, XL	N	FDA approved product preferred. FDA approved manufacturer preferred. Nitrile preferred. May substitute vinyl or polypropylene. Does not require FDA approval. https://www.fda.gov/media/136540/download	FDA products (preferred) FDA manufacturer ASTM D6319-19 (preferred nitrile std) ASTM D5250-19 (preferred vinyl std) ASTM D6977-19 (preferred polychloroprene std) ISO 374-5 ASTM F2878	May 29 2020

FDA code	MMC Item code	Product type	Size	Sterile	Specs	Certification	DOH approval date
OEA		Gown, level 2	S/M, L, XL, XXL	N	Meets FDA approval, any listed certification or description states liquid resistant. Does not need FDA approval. Must be listed as AAMI Level 2 or be described for "low risk exposure" (for example during blood draw, suturing, in the Intensive Care Unit (ICU), or a pathology lab). Note AAMI protection level 2 on purchase/tracking form. https://www.fda.gov/media/136540/download https://www.cdc.gov/niosh/nppt/topics/protectiveclothing/	FDA products (preferred) FDA manufacturer ANSI/AAMI PB70 (level 2), AATCC 42 ($\leq 1.0g$)/ AATCC 127 ($\geq 20cm$) ISO 22610 (2.8 or higher) EN 13795 (standard performance ≥ 3 rating/ $\geq 3.5kPa$) EN 14126 (≥ 3 rating/ $\geq 3.5kPa$) GB 19802 (≥ 3 rating/ $\geq 3.5kPa$) GB/T 4745 (≥ 3 rating/ $\geq 3.5kPa$) GB/T 4744 (≥ 3 rating/ $\geq 3.5kPa$)	May 29 2020
		Hand Sanitizer (70% or greater alcohol)	24 oz. bottles (Preferred)		70% or greater alcohol preferred. Does not need FDA approval. S Companies manufacturing alcohol for hand sanitizer have to register with FDA. Searchable FDA drug establishment https://www.fda.gov/media/136289/download https://www.fda.gov/media/136390/download	FDA products FDA manufacturer	April 1 2020
		Hand Sanitizer Refill	1200 ml or 2000 ml		70% or greater alcohol preferred. Does not need FDA approval. S Companies manufacturing alcohol for hand sanitizer have to register with FDA. Searchable FDA drug establishment https://www.fda.gov/media/136289/download https://www.fda.gov/media/136390/download	FDA products FDA manufacturer	April 1 2020

FDA code	MMC Item code	Product type	Size	Sterile	Specs	Certification	DOH approval date
QKR		Mask, face		N	Disposable face mask. Ear loop or string ties. Not rated for surgery, not rated for liquid protection, not rated for flammable or high heat. Prefer FDA approved, but not required. https://www.fda.gov/media/136449/download https://www.fda.gov/media/137121/download	FDA products FDA manufacturer	April 8 2020
		Mask, face cloth		N	Reusable cotton or cotton poly blend cloth face mask. Must be at least 180 thread count. Sewn in string ties preferred; sewn in ear loops acceptable. All edges must be stitched to prevent fraying. No added copper, silver or other antimicrobial agents. Prefer assortment of sizes. Must have double layer fabric. Prefer pocket to add disposable filter material. Product label must comply with current FDA guidelines for product that will be used in healthcare settings. Product for consumer use does not need to follow FDA EUA. Manufacturer should review FDA Enforcement Policy and EUA for further information. https://www.fda.gov/media/137121/download https://www.fda.gov/media/136449/download	≥180 thread count cotton or cotton poly blend double layer fabric. Non stretch fabric.	July 8 2020

FDA code	MMC Item code	Product type	Size	Sterile	Specs	Certification	DOH approval date
		Respirator, KN95 disposable	OS		KN95 respirator certified to GB2626-2006 listed on Appendix A,	GB2626 Importing Chinese filtering face piece https://www.fda.gov/media/136664/download Appendix A – approved Chinese manf https://www.fda.gov/media/136663/download Enforcement policy https://www.fda.gov/media/136449/download	May 29 2020
FLL		Thermometer non contact			range between 96.0-104°F and accurate to within 0.2°F https://www.fda.gov/media/136698/download	ASTM 11104-98; ASTM 1965-98; ASTM E112; ISO 80601-2-56; ASTM E1061-01;	April 8 2020